



RELATIVE HUMIDITY MEASUREMENT

**INTERLABORATORY COMPARISON
APLAC M023**

Final Report
April 2010

ACKNOWLEDGEMENTS

The APLAC Secretariat wishes to acknowledge gratefully the financial support provided by STANDARDS MALAYSIA and also to APLAC members for their participation and cooperation.

© COPYRIGHT ASIA PACIFIC LABORATORY ACCREDITATION COOPERATION, 2009

The contents of this report shall not be reproduced without written permission of the Secretary of APLAC.

CONTENTS

Section	Page
1. Introduction	3
2. Aim of the Programme	3
3. Design of the Programme	3
4. Participants	4
5. Circulation Delays	4
6. Reporting by Participants	5
7. Reference Values and Stability of the Artefacts	5
8. Interim Reports	6
9. Summary of Result	6
10. Performance of Participants	23
11. Uncertainties of Measurement	25
12. Corrective Actions	25
13. References	25
 Appendices	
A Instructions to Laboratories and Result Sheets	
B Circulation Schedule	
C Reference value	
D Performance of Artefacts before and after each calibration	
E Example of Uncertainty budget table	

1. INTRODUCTION

This report summarises the results of an interlaboratory comparison on relative humidity calibration by laboratories accredited by bodies that are members of the Asia Pacific Laboratory Accreditation Cooperation (APLAC). The identification code number of this programme is M023. This programme was coordinated by the STANDARDS MALAYSIA.

2. AIM OF THE PROGRAM

One of the APLAC's main aims is to build up and maintain mutual confidence in the technical competence of APLAC members. Interlaboratory comparisons provide objective evidence that laboratories are competent and that they can achieve the level of accuracy for which they are accredited. They also provide means for improving the quality and performance of laboratories throughout the region.

APLAC M023 relative humidity programme was designed to test the participant's ability to perform calibration of a relative humidity using a relative humidity and temperature transmitter over a specified range and to obtain values within their accredited uncertainties of measurement. Secondary aim is to assess the participant's ability to estimate their uncertainty of measurement.

3. DESIGN OF THE PROGRAMME

This interlaboratory comparison involved the circulation of relative humidity and temperature transmitter with digital indicator around two schemes of participating laboratories, designated as Scheme A and Scheme B. The measurement artefacts comprised of two Vaisala relative humidity and temperature transmitter with digital indicator. The artefacts were identified as Artefact A and Artefact B and were circulated to Scheme A and Scheme B respectively.

National Metrology Laboratory (NML) of SIRIM, Malaysia acted as the Reference Laboratory, deriving values from a series of calibrations of the measurement artefacts at the beginning and the end of the program and at the intervals throughout.

The selected instrument needed to have sufficient resolution, repeatability and stability to be suitable for laboratories.

Participants were required to carry out calibration from 10 %rh to 90 %rh at nominal temperature of 23 °C. Participants were allowed to use their own measurement methods. A copy of the *Instructions to Laboratories* appears in Appendix A.

4. PARTICIPANTS

21 laboratories from 10 accreditation bodies participated in the programme. These laboratories are accredited for various levels of accuracy as defined in their scope of accreditation which lists their least uncertainties of measurement. During the circulation of the programme three laboratories withdrawn their participation.

In order to keep with the schedule, all the accreditation bodies were asked to strictly follow the circulation period given to all laboratories before starting the programme. However the circulation delayed for a few laboratories due to custom regulation for that particular country. Appendix B illustrated the circulation schedule for each laboratories.

To maintain confidentiality, participating laboratories were assigned random Lab Code numbers and these numbers are used in the report.

5. CIRCULATION DELAYS

There were delays in the circulation of artefacts. Most delays were only for a few weeks. The delays were due to the custom regulation in various countries.

6. REPORTING BY PARTICIPANTS

Laboratories were requested to report their test results on a pro-forma *Results Sheet* (refer to Appendix A). The participating laboratories shall report the calibration uncertainties at 95% confidence levels.

7. REFERENCE VALUES AND STABILITY OF THE ARTEFACTS

7.1 *Assignment of Reference Values*

Before and after each round of circulation, the reference laboratory carried out a series of calibration of the artefacts. The reference values obtained from the calibration of the artefacts are in terms of correction to the readings at nominal value. Assignment of reference values for each laboratory is the average of readings before and after each round of circulation. Each loop had been assigned with different reference value. The results obtained by the reference laboratory appear in Appendix C.

The calibrations performed by the reference laboratory before and after the circulation showed that drift occurred for both artefacts for all ranges. Referring to APLAC PT001 *Calibration Interlaboratory*, where drift had occurred significantly during the circulation, different reference value may be specified for different laboratories. Therefore no laboratory is unfairly disadvantaged. The transportation drift value was included in the uncertainty calculation for the reference value. Appendix E explained how and where drift was included in uncertainty calculation. The method used by the reference laboratory for uncertainty calculation was based on the method described in the ISO /IEC Guide 98-3, *Uncertainty of Measurement*.

7.2 *Performance of Artefacts*

Laboratories were asked to carry out measurement at 50 %rh before start calibration and after completed calibration in order to ensure that artefacts perform satisfactorily for the entire programme. Appendix D provides results of the measurement fluctuated within ± 2.0 %rh at 50 %rh. Therefore both artefact performed well during the programme.

8. INTERIM REPORTS

Upon receiving the calibration reports from the participating laboratories and completion of measurement of the reference value, interim reports were issued to the participating laboratories.

9. SUMMARY OF RESULTS

In accordance with the international practice, measurement performance has been assessed on the basis of an E_n ratio for each measurement. The E_n ratios are calculated using a standard statistical technique for comparing values and are derived from the following expression:

$$E_n = \frac{LAB - REF}{\sqrt{U_{LAB}^2 + U_{REF}^2}}$$

where:

LAB is the correction at nominal value measured by the participating laboratory.

REF is the average of correction at nominal value determined from measured correction value made before and after the circulations round by the reference laboratory.

U_{LAB} is the expanded uncertainties of the correction provided by the participating laboratories.

U_{REF} is the expanded uncertainties of the correction evaluated by the reference laboratory and the transportation drift.

For the results to be acceptable, values of $|E_n| \leq 1$ are required. Values of $|E_n| > 1$ require investigation and corrective action by both the laboratory concerned and its accreditation body.

A summary of results for every participating laboratories are summarized in Tables 1 to 2 along with the $|E_n|$ values. The graphs in Figures 1 to 10 illustrate the correction values corrected to nominal values reported by each laboratory together with its associated uncertainty.

Table 1 – Summary of Reported Results Artefact A

Relative humidity: 10 %rh at 23 °C

Lab Code	Unit	214	217	237	239	253	273	278	296			
Correction at nominal for <i>Lab</i>	%rh	0.19	0.33	-0.57	0.13	-0.21	-0.20	-0.65	0.13			
Correction at nominal for <i>Ref</i>	%rh	0.36	0.86	0.36	0.36	-0.13	-0.13	-0.13	-0.13			
<i>LAB-REF</i>		-0.17	-0.52	-0.93	-0.23	-0.08	-0.07	-0.51	0.26			
U_{Lab}	%rh	0.4	1.7	0.2	0.2	2.1	1.1	2.3	1.7			
U_{Ref}	%rh	1.1	1.6	1.1	1.1	2.0	2.0	2.0	2.0			
Coverage Factor: <i>k</i>	-	2	2	2	2	2	2	2	2			
$ E_n $	-	0.14	0.22	0.81	0.20	0.03	0.03	0.17	0.10			

Relative humidity: 30 %rh at 23 °C

Lab Code	Unit	214	215	217	237	239	251	253	257	273	278	293	296
Correction at nominal for <i>Lab</i>	%rh	0.78	1.13	0.60	0.59	0.84	1.3	0.54	0.21	0.85	0.88	1.90	0.06
Correction at nominal for <i>Ref</i>	%rh	0.92	0.79	0.79	0.92	0.92	0.8	0.21	0.21	0.21	0.21	0.21	0.21
<i>LAB-REF</i>		-0.14	0.34	-0.19	-0.34	-0.08	0.54	0.32	-0.00	0.63	0.66	1.68	-0.15
<i>U_{Lab}</i>	%rh	0.7	1.3	2.0	0.2	0.3	3.2	2.1	3.5	1.1	2.3	2.3	1.7
<i>U_{Ref}</i>	%rh	1.2	1.6	1.6	1.2	1.2	1.6	2.0	2.0	2.0	2.0	2.0	2.0
Coverage Factor: <i>k</i>	-	2	2	2	2	2	2	2	2	2	2	2	2
$ E_n $	-	0.10	0.17	0.07	0.28	0.07	0.15	0.11	0.00	0.28	0.22	0.55	0.06

Relative humidity: 50 %rh at 23 °C

Lab Code	Unit	214	215	217	237	239	251	253	257	273	278	293	296
Correction at nominal for L_{Lab}	%rh	1.94	2.57	1.72	1.62	1.71	2.4	1.55	1.03	2.40	2.45	3.41	0.35
Correction at nominal for Ref	%rh	1.86	1.73	1.73	1.86	1.86	1.73	1.27	1.27	1.27	1.27	1.27	1.27
L_{AB-REF}		0.08	0.84	-0.02	-0.24	-0.15	0.69	0.28	-0.24	1.13	1.18	2.14	-0.92
U_{Lab}	%rh	1.1	1.5	2.3	0.2	0.5	2.9	2.1	3.5	1.1	2.3	2.3	1.7
U_{Ref}	%rh	1.3	1.7	1.7	1.3	1.3	1.7	2.1	2.1	2.1	2.1	2.1	2.1
Coverage Factor: k	-	2	2	2	2	2	2	2	2	2	2	2	2
$ E_n $	-	0.05	0.37	0.01	0.19	0.11	0.21	0.10	0.06	0.48	0.38	0.69	0.34

Relative humidity: 70 %rh at 23 °C

Lab Code	Unit	214	215	217	237	239	251	253	257	273	278	293	296
Correction at nominal for L_{lab}	%rh	3.04	3.33	2.96	2.24	2.50	3.4	2.32	1.72	3.82	3.82	5.52	0.93
Correction at nominal for Ref	%rh	2.73	2.87	2.87	2.73	2.73	2.9	2.32	2.32	2.32	2.32	2.32	2.32
L_{AB-REF}		0.30	0.46	0.10	-0.49	-0.24	0.57	0.00	-0.60	1.50	1.50	3.20	-1.40
U_{Lab}	%rh	1.5	1.7	2.7	0.2	0.7	2.9	2.5	3.5	1.1	2.3	2.3	2.0
U_{Ref}	%rh	1.4	1.8	1.8	1.4	1.4	1.8	2.2	2.2	2.2	2.2	2.2	2.2
Coverage Factor: k	-	2	2	2	2	2	2	2	2	2	2	2	2
$ E_n $	-	0.15	0.19	0.03	0.34	0.15	0.17	0.00	0.15	0.61	0.47	1.01	0.47

Relative humidity: 90 %rh at 23 °C

Lab Code	Unit	214	215	217	237	239	251	253	257	273	278	293	296
Correction at nominal for L_{lab}	%rh	3.44	3.41	3.62	2.21	2.98	4.4	2.32	1.74	4.48	4.73	2.81	1.72
Correction at nominal for Ref	%rh	3.12	3.38	3.38	3.12	3.12	3.4	2.64	2.64	2.64	2.64	2.64	2.64
$LAB-REF$		0.32	0.04	0.24	-0.91	-0.14	0.98	-0.32	-0.90	1.84	2.09	0.17	-0.92
U_{Lab}	%rh	1.8	2.0	3.0	0.2	0.9	2.9	2.5	3.5	1.1	2.3	2.3	2.6
U_{Ref}	%rh	1.6	1.9	1.9	1.6	1.6	1.9	2.4	2.4	2.4	2.4	2.4	2.4
Coverage Factor: k	-	2	2	2	2	2	2	2	2	2	2	2	2
$ E_n $	-	0.14	0.01	0.07	0.57	0.08	0.29	0.09	0.21	0.70	0.63	0.05	0.26

Table 2 – Summary of Reported Results Artefact B

Relative humidity: 10 %rh at 23 °C

Lab Code	Unit	224	260	266	275	279			
Correction at nominal for <i>Lab</i>	%rh	0.7	0.23	0.23	0.4	0.2			
Correction at nominal for <i>Ref</i>	%rh	-0.1	-0.34	-0.34	-0.1	-0.1			
<i>LAB-REF</i>		0.75	0.57	0.58	0.51	0.34			
U_{Lab}	%rh	0.5	0.4	0.4	1.2	0.3			
U_{Ref}	%rh	1.4	1.2	1.2	1.4	1.4			
Coverage Factor: k	-	2	2	2	2	2			
$ E_n $	-	0.50	0.45	0.46	0.28	0.24			

Relative humidity: 30 %rh at 23 °C

Lab Code	Unit	224	260	263	266	267	271	275	279
Correction at nominal for Lab	%rh	1.6	0.60	1.4	0.56	-0.11	0.8	1.1	1.1
Correction at nominal for Ref	%rh	1.0	0.40	0.9	0.40	0.40	0.9	1.0	1.0
$LAB-REF$		0.67	0.21	0.50	0.16	-0.50	-0.13	0.15	0.11
U_{Lab}	%rh	0.5	0.5	4.9	1.5	2.3	2.3	1.2	0.3
U_{Ref}	%rh	1.5	1.3	1.7	1.3	1.3	1.7	1.5	1.5
Coverage Factor: k	-	2	2	2	2	2	2	2	2
$ E_n $	-	0.42	0.15	0.10	0.08	0.19	0.05	0.08	0.07

Relative humidity: 50 %rh at 23 °C

Lab Code	Unit	224	260	261	263	266	267	271	275	279
Correction at nominal for <i>Lab</i>	%rh	2.9	1.85	1.13	-1.4	1.30	1.45	2.3	2.3	2.3
Correction at nominal for <i>Ref</i>	%rh	2.2	1.42	1.42	2.3	1.42	1.42	2.3	2.2	2.2
<i>LAB-REF</i>		0.76	0.43	-0.29	-3.7	-0.12	0.03	-0.04	0.12	0.14
U_{Lab}	%rh	0.5	0.6	2.4	4.9	2.0	2.4	2.3	1.2	0.3
U_{Ref}	%rh	1.6	1.4	1.4	1.7	1.4	1.4	1.7	1.6	1.6
Coverage Factor: <i>k</i>	-	2	2	2	2	2	2	2	2	2
$ E_n $	-	0.45	0.28	0.10	0.7	0.05	0.01	0.02	0.06	0.08

Relative humidity: 70 %rh at 23 °C

Lab Code	Unit	224	260	261	263	266	267	271	275	279
Correction at nominal for <i>Lab</i>	%rh	3.9	2.92	1.83	-1.3	2.94	2.23	3.5	3.4	3.4
Correction at nominal for <i>Ref</i>	%rh	3.2	2.38	2.38	3.5	2.38	2.38	3.5	3.2	3.2
<i>LAB-REF</i>		0.73	0.54	-0.55	-4.9	0.56	-0.15	0.00	0.19	0.21
U_{Lab}	%rh	0.5	0.8	2.4	4.9	2.8	2.4	2.3	1.2	0.3
U_{Ref}	%rh	1.7	1.6	1.6	1.8	1.6	1.6	1.8	1.7	1.7
Coverage Factor: <i>k</i>	-	2	2	2	2	2	2	2	2	2
$ E_n $	-	0.41	0.30	0.19	0.9	0.17	0.05	0.00	0.09	0.12

Relative humidity: 90 %rh at 23 °C

Lab Code	Unit	224	260	263	266	267	271	275	279
Correction at nominal for <i>Lab</i>	%rh	4.0	2.76	-1.8	5.98	3.11	4.6	3.8	3.8
Correction at nominal for <i>Ref</i>	%rh	3.7	2.95	4.1	2.95	2.95	4.1	3.7	3.7
<i>LAB-REF</i>		0.28	-0.20	-5.9	3.02	0.16	0.42	0.13	0.08
U_{Lab}	%rh	1.0	0.9	4.9	4.2	2.6	2.3	1.2	0.3
U_{Ref}	%rh	1.9	1.8	1.9	1.8	1.8	1.9	1.9	1.9
Coverage Factor: k	-	2	2	2	2	2	2	2	2
$ E_n $	-	0.13	0.10	1.1	0.66	0.05	0.14	0.06	0.04

Corrections at nominal values with its associated uncertainty for Artefact A: C4510044




Figure 1: Calibration at 10 %rh at 23 °C




Figure 2: Calibration at 30 %rh at 23 °C




Figure 3: Calibration at 50 %rh at 23 °C




Figure 4: Calibration at 70 %rh at 23 °C




Figure 5: Calibration at 90 %rh at 23 °C

Corrections at nominal values with its associated uncertainty for Artefact B : C4510045




Figure 6: Calibration at 10 %rh at 23 °C




Figure 7: Calibration at 30 %rh at 23 °C




Figure 8: Calibration at 50 %rh at 23 °C




Figure 9: Calibration at 70 %rh at 23 °C




Figure 10: Calibration at 90 %rh at 23 °C

10. PERFORMANCE OF PARTICIPANTS

Tables 1 and 2 for the $|E_n|$ ratios show an overall good performance by the majority of the participating laboratories. The following is the summary of performance of participating laboratories.

10.1 *Circulation of Scheme A*

Twelve (12) laboratories registered in measurement for Scheme A. For measurement at 10 %rh, 30 %rh, 50 %rh, and 90 %rh, the performance of all participating laboratories was excellent as indicated by the fact that no laboratory had an $|E_n|$ value greater than 1. For measurement at 70 %rh, Lab Code 293 had one result with an $|E_n|$ values exceeded 1.

Only eight (8) laboratories were able to perform measurement at 10 %rh. All twelve (12) laboratories were able to perform measurement from 30 %rh to 90 % rh.

10.2 *Circulation of Scheme B*

Nine (9) laboratories registered in measurement for Scheme A. For measurement at 10 %rh, 30 %rh, 50 %rh, and 70 %rh, the performance of all participating laboratories was excellent as indicated by the fact that no laboratory had an $|E_n|$ value greater than 1. For measurement at 90 %rh, Lab Code 263 had one result with an $|E_n|$ values exceeded 1.

Only five (5) laboratories were able to perform measurement at 10 %rh. Eight (8) laboratories were able to performed measurement at 30 %rh. All nine (9) laboratories were able to perform measurement from 50 %rh to 70 % rh. Finally eight (8) laboratories were able to performed measurement at 90 %rh.

As shown in Figure 11 for 95 reported results, 98% were in agreement with the reference laboratory and 2% were outliers.




Figure 11 – Summary of 95 Reported Results

Figure 12 shows that 91% of the participating laboratories obtained an excellent performance with an $|E_n|$ value of less than unity for all the measurements, 9% of the participating laboratories each had one result with an $|E_n|$ value exceeded 1.




Figure 12 - Summary of 21 Participating Laboratories

11. UNCERTAINTIES OF MEASUREMENT

Uncertainty calculations were generally adequate and covered the major sources of uncertainty. Uncertainty calculation for reference value seems to be quite large compared to a few participating laboratories. This is due to the effect of transportation drift for each artefact. Five (5) laboratories reported uncertainty value of less than 1 %rh. Most of the laboratories reported uncertainty values between 1.0 %rh to 3.0 %rh. However, Lab code 257 and 263 reported large values of expanded uncertainty (more than 3.0 %rh) compare to others laboratories.

12. CORRECTIVE ACTIONS

Details of corrective action are not included in this report. The respective accreditation bodies (ABs) have to ensure that any outlier under their purview has the problem rectified and procedures to prevent recurrence have been put in place.

13. REFERENCES

1. APLAC PT001 Calibration Interlaboratory Comparisons Issue No. 5
2. ISO /IEC Guide 98-3, Uncertainty of Measurement – Part 3: Guide to the Expression of Uncertainty in Measurement.
4. Guide to calculate uncertainty in humidity measurement by NPL, UK.
5. ISO/IEC 17025 General requirements for the competence of testing and calibration laboratories.



Asia Pacific Laboratory Accreditation Cooperation

APPENDIX A

APLAC M023 RELATIVE HUMIDITY MEASUREMENT PROFICIENCY TESTING PROGRAMME

**CALIBRATION OF
RELATIVE HUMIDITY 10 %rh TO 90 %rh AT NOMINAL
TEMPERATURE OF 23 °C**

MEASUREMENT INSTRUCTIONS

October 2008

APLAC PROFICIENCY TESTING PROGRAMME M023

APLAC M023 CALIBRATION OF RELATIVE HUMIDITY 10 %rh TO 90 %rh AT NOMINAL TEMPERATURE OF 23 °C

1. Introduction

The purpose of APLAC M023 Relative Humidity Measurement Proficiency Testing Programme is to compare the calibration results of accredited laboratories within APLAC in the field of Relative Humidity measurement. Two different relative humidity and temperature sensor of the same model will be used in a circulation, among the accredited laboratories of the APLAC members.

2. Coordinator

The coordinator of this PT Programme is the Department of Standards Malaysia. The complete information is as follows:

STANDARDS MALAYSIA
Century Square,
Level 1 & 2, Block 2300,
Jalan Usahawan,
63000 Cyberjaya, Selangor
MALAYSIA

Contact person for

Department of Standards Malaysia

Mr. Ahmad Razif Bin Abd. Rahim

Tel : +60 3 8319 1400

Fax : +60 3 8318 9339

E-mail : razif@standardsmalaysia.gov.my

3. Reference Laboratory

The reference values for the Relative Humidity measurement will be provided by the National Metrology Laboratory, SIRIM Berhad, (NML-SIRIM), Malaysia. Complete information on NML-SIRIM is as follows:

National Metrology Laboratory
SIRIM Berhad
Lot PT 4803, Bandar Baru Salak Tinggi
43900 Sepang
MALAYSIA

Contact person for NML-SIRIM

Ms. Faridah Hussain
Tel: +60 3 8778 1735
Fax : +60 3 8778 1661
E-mail: faridah_hussain@sirim.my

4. Artefact

Each accredited laboratories will be provided with one relative humidity and temperature transmitter with digital indicator. Each device comes with one sensor, each measuring temperature and relative humidity. It will be shipped together with an operation manual. Details of the artefact are as below:

Artefact	Measurement Temperature (°C)	Manufacturer	Model	Serial Number
A	23 °C	Vaisala	HMT330	C4510044
B	23 °C	Vaisala	HMT330	C4510045

5. Calibration Procedure

During the calibration, the artefacts should be supplied with $240\text{ V} \pm 10\%$, 50 Hz. Measurements will be performed using the Laboratory's own procedures, within the following constraints:

All laboratories should refer to the operating manuals for instructions and precautions for using the artefacts. These instructions shall be read in full and if in doubt on any of the aspects, the laboratory shall contact the reference laboratory. In the case of any unexpected instrument failure at a participant laboratory, the reference laboratory should be informed in order to revise the schedule, if necessary, as early as possible.

- Measurements shall be performed for increasing values of relative humidity.

- No adjustments shall be performed on the artefacts.
- The artefact used in this comparison must **not be** modified, adjusted, or used for any purpose other than described in this document, nor given to any party other than the participants in the comparison.

5.1 Measurement at 23 °C

Equipment:

- Artefact A or B
 - Laboratory relative humidity and temperature standard
 - Laboratory relative humidity and temperature calibration chamber
- The artefact shall be left in the nominal ambient conditions of the laboratory for a minimum period of 3 hours before taking any measurement.
 - Place the sensor of the artefact in a relative humidity and temperature calibration chamber together with the laboratory relative humidity and temperature standards. Refer to Set-up diagram in Figure 1.




Figure 1: Schematic diagram set-up for calibration

- For checking purposes of the artefact before calibration, set the relative humidity and temperature calibration chamber to **50 %rh** at **23 °C**. Take the measured data for standard used and the artefact using **Appendix F1**.
- Then perform measurement at the nominal value within the Laboratory's scope of accreditation at nominal temperature of **23 °C**.

5. Nominal values to be measured: **10 %rh, 30 %rh, 50 %rh, 70 %rh** and **90 %rh**. Measurement should be made in **rising order** of relative humidity.
6. At each generated condition a time of not less than 60 minutes should be allowed for the relative humidity to equilibrate. A set of **10** readings at **1 minute interval** should be recorded from the laboratory standard and artefact.
7. Once again after completed all measurement points, set the relative humidity and temperature calibration chamber to **50 %rh** at **23 °C**. Take the measured data for standard used and the artefact using **Appendix F2**.

6. Environmental Conditions

Assure that the calibration ambient temperature is stable and within 18 °C to 28 °C and the ambient relative humidity is less than 80 %.

The actual environmental conditions during calibration should be reported.

7. Reporting of calibration results

Please refer to the table below for the details of results submission.

Document	When to be submitted	Submitted to whom
Completed proforma result sheets (Appendix F1-7)	Within 2 weeks after completing the calibration	Participating Laboratory Accreditation Body (AB) for forwarding to STANDARDS MALAYSIA
Completed laboratory's budgets for estimating the uncertainty of measurement. Please refer to Appendix G for the template of estimating uncertainty of measurement.		

8. Analysis of results

The reference laboratory will analyze the results and produce interim and final reports in consultation with the STANDARDS MALAYSIA.

Appendix A

Instructions to Accreditation Bodies

APLAC PROFICIENCY TESTING PROGRAMME M023 – RELATIVE HUMIDITY MEASUREMENT

1. EQUIPMENT

On receipt, unpack the audit pack and inspect the artefact for any defects. *List other specific checks.*

Complete the attached "*RECEIPT FORM*" and fax to STANDARDS MALAYSIA.

2. TRANSPORT

The artefacts can be carried by hand, car or by plane, whatever means of transportation considered as safe.

3. CUSTOMS

NO ATA Carnet is provided with these artefacts. A *Declaration to Customs Officials and Shipping Agents* is enclosed in the plastic envelope attached to the outside of the box. Please make sure it is in place before dispatching to the next country otherwise it will not pass through their Customs authorities. Please attach 3 extra photocopies of the *Declaration*.

4. MEASUREMENTS TO BE CARRIED OUT

Please refer to *MEASUREMENT INSTRUCTIONS* enclosed. If necessary, these should be translated into the appropriate language by the accreditation body.


5. CIRCULATION SCHEME

At the beginning, middle and at the end of the circulation scheme, measurements will be performed by the reference laboratory, National Metrology Laboratory, Malaysia. Two (2) units of artefact; A and B will be used to cover two (2) parallel scheme ; scheme A and scheme B.


Each laboratory will be given 1(one) weeks to perform measurement. It is the responsibility of the accreditation body to make sure that the artefact is dispatched to the next participant by the date specified. One week have been allowed for each international transport and customs clearance. If a delay occurs with one laboratory, STANDARD MALAYSIA shall be informed. When ready for dispatch, fill in the attached *DISPATCH FORM* and fax to the next participant and to STANDARD MALAYSIA.

CIRCULATION SCHEME FOR TRAVELLING ARTEFACT

SCHEME A ARTEFACT A



SCHEME B
ARTEFACT B



6. DOCUMENTS TO BE SUBMITTED

Within two week of the completion of the measurements, participating laboratories are required to fax or send the *Result Sheet* together with their formal calibration report to their accreditation body. The accreditation body is required to ensure that the *Result Sheet* has been filled in correctly and completely. The accreditation body should provide details of any problems that their laboratories had. The accreditation body is required to fax the documentation to the organising body within two weeks of dispatching the artefact.

On receipt of the information, the organising body for this APLAC proficiency testing programme will prepare interim reports on each laboratory which are then sent back to the national accreditation body. A final report will be issued at the end of the program with each laboratory only identified by a **confidential** code number.

7. GENERAL INFORMATION

For general queries, please contact:

STANDARDS MALAYSIA
Century Square,
Level 1 & 2, Block 2300,
Jalan Usahawan,
63000 Cyberjaya, Selangor
MALAYSIA

Attention: Mr. Ahmad Razif Bin Abd. Rahim
Fax : +60 3 8318 9339
E-mail : razif@standardsmalaysia.gov.my

Appendix B Receipt Form

APLAC PROFICIENCY TESTING PROGRAMME M023 – RELATIVE HUMIDITY MEASUREMENT

In order to monitor the progress of the PT Programme, we kindly ask each accreditation body, on receipt of the artefact, to fill in this RECEIPT FORM and fax it to:

STANDARDS MALAYSIA
Century Square,
Level 1 & 2, Block 2300,
Jalan Usahawan,
63000 Cyberjaya, Selangor
MALAYSIA

Attention: Mr. Ahmad Razif Bin Abd. Rahim
Fax : +60 3 8318 9339
E-mail : razif@standardsmalaysia.gov.my

Thank you in advance for your cooperation.

The M023 artefact, serial number _____ was received on: _____ (date)

After inspection, are the contents damaged? _____
(yes/no)

If yes, is this serious? _____
(yes/no)

Are the contents still suitable for use? _____
(yes/no)

Was there a "*Declaration to Customs Officials and Shipping Agents*" enclosed in the plastic envelope attached to the outside of the case?

(yes/no)

Remarks: _____

Participating Accreditation Body: _____

Contact Person: _____ Fax: _____

e-mail: _____ Tel: _____

Appendix C Dispatch Form

APLAC PROFICIENCY TESTING PROGRAMME M023 – RELATIVE HUMIDITY MEASUREMENT

In order to monitor the progress of the PT Programme, we kindly ask each accreditation body, on dispatch of the artefact, to fill in this DISPATCH FORM and fax it to:

STANDARDS MALAYSIA
Century Square,
Level 1 & 2, Block 2300,
Jalan Usahawan,
63000 Cyberjaya, Selangor
MALAYSIA

Attention: Mr. Ahmad Razif Bin Abd. Rahim
Fax : +60 3 8318 9339
E-mail : razif@standardsmalaysia.gov.my

and **also** fax it to the next participating accreditation body :-

Please ensure that the "*Declaration to Customs Officials and Shipping Agents*" is attached to the outside of the case. Thank you in advance for your cooperation.

The M023 artefact, serial number _____ was dispatched on: _____
(date)

The artefact has been inspected after return from our
laboratories and were found to be in good condition. _____ (yes/no)
Please give details of any problems:-

Shipping agent: _____ Phone: _____ Fax: _____

Airwaybill no. (or consignment note no.): _____

Your accreditation body: _____

Contact person: _____ Fax: _____

e-mail: _____ Tel: _____

Appendix D

Instructions to Laboratories

APLAC PROFICIENCY TESTING PROGRAMME M023 – RELATIVE HUMIDITY MEASUREMENT

1. Equipment

The contents of the audit pack are as follows:

Item No.	Unit	Description
1.	One (1)	Temperature and Humidity Transmitter, Brand Vaisala, Model HMT330, S/No.C4510044 or C4510045
2.	One (1)	User's Guide * Please put back the User's Guide in the transportation box before sending to another laboratory



Figure 2: Photo of the artefact

2. On Receipt

On receipt, unpack the audit pack and inspect the artefact for any defects.

Read the following instructions before starting the measurement.

1. Please handle artefact with care.
2. Do not tamper with or adjust the test artefact.
3. Notify the local AB of any damage or erratic operation of the test artefact as soon as possible.

3. Measurement Considerations for the Artefact

Power up the artefact using 240 VAC, 50 Hz supply from the mains.

Refer to **Measurement Instructions** for the details of measurement to be performed.

4. Completion of measurements

Upon completion of required measurement, the laboratories are required to inform the respective accreditation body in the country.

5. Documents to be submitted

Within two weeks of the completion of the measurements, participating laboratories are required to send the attached *Result Sheet* (**Appendix A, B, C, D, E, F1 to F7, G**) and the calibration report to the respective accreditation body. Laboratories are strongly advised to make a copy of the *Result Sheet* for record keeping and future reference purposes.

In addition to the above, participating laboratories are required to submit uncertainty budgets which are prepared in accordance with the *ISO Guide to the Expression of Uncertainty in Measurement*. See **Appendix G** for the template of the uncertainty budget table. **Participating laboratories are required to prepare uncertainty budget table for each calibration point.**

A final report will be issued at the end of the program with each laboratory which is identified by a confidential code number.

6. General Information

For general queries, please contact your accreditation body. Additional information may be obtained from :

STANDARDS MALAYSIA
Century Square,
Level 1 & 2, Block 2300,
Jalan Usahawan,
63000 Cyberjaya, Selangor
MALAYSIA

Attention: Mr. Ahmad Razif Bin Abd. Rahim
Tel : +60 3 8319 1400
Fax : +60 3 8318 9339
E-mail : razif@standardsmalaysia.gov.my

Appendix E

LABORATORY DETAILS

M023 ☐☐☐☐
For official use only

Name of laboratory	
Measurement Date	From _____ to _____
Mains Power Supply	_____ VAC $\pm$ _____ VAC, _____ Hz
Ambient Temperature (°C)	_____ °C $\pm$ _____ °C
Ambient Relative Humidity (%)	_____ % $\pm$ _____ %
Laboratory Reference Standard	Description: Maker: Model: Serial no.: Calibration due date :
Laboratory Humidity calibration chamber	Description: Maker: Model: Serial no.:

Appendix F1

RESULT SHEET

Measurement point: 50 %rh at 23 °C

BEFORE CALIBRATION

	Corrected value of Laboratory Standard (%rh)	Corrected value of Laboratory Standard t, (°C)	Indicated value of Artefact _____ (%rh)	Indicated value of Artefact _____ t, (°C)
1				
2				
3				
4				
5				
6				
7				
8				
9				
10				
Average				
Std dev.				

Name of the Laboratory : _____

Date of Measurement : _____

Environment of the Laboratory : Temperature : (____ ± ____) °C

Humidity : (____ ± ____) %rh

Appendix F2

RESULT SHEET

Measurement point: 50 %rh at 23 °C

AFTER CALIBRATION

	Corrected value of Laboratory Standard (%rh)	Corrected value of Laboratory Standard t, (°C)	Indicated value of Artefact _____ (%rh)	Indicated value of Artefact _____ t, (°C)
1				
2				
3				
4				
5				
6				
7				
8				
9				
10				
Average				
Std dev.				

Name of the Laboratory : _____

Date of Measurement : _____

Environment of the Laboratory : Temperature : (____ ± ____) °C

Humidity : (____ ± ____) %rh

Appendix F3

RESULT SHEET

ARTEFACT _____

Measurement point: 10 %rh at 23 °C

Nominal value 10%	Corrected value of Laboratory Standard (%rh)	Corrected value of Laboratory Standard t, (°C)	Indicated value of Artefact (%rh)	Indicated value of Artefact t, (°C)
1				
2				
3				
4				
5				
6				
7				
8				
9				
10				
Average				
Std dev				

Calibrated By:

Checked By:

Date:

Date:

Appendix F4

RESULT SHEET

ARTEFACT _____

Measurement point: 30 %rh at 23 °C

Nominal value 30%	Corrected value of Laboratory Standard (%rh)	Corrected value of Laboratory Standard t, (°C)	Indicated value of Artefact (%rh)	Indicated value of Artefact t, (°C)
1				
2				
3				
4				
5				
6				
7				
8				
9				
10				
Average				
Std dev				

Calibrated By:

Checked By:

Date:

Date:

Appendix F5

RESULT SHEET

ARTEFACT _____

Measurement point: 50 %rh at 23 °C

Nominal value 50%	Corrected value of Laboratory Standard (%rh)	Corrected value of Laboratory Standard t, (°C)	Indicated value of Artefact (%rh)	Indicated value of Artefact t, (°C)
1				
2				
3				
4				
5				
6				
7				
8				
9				
10				
Average				
Std dev				

Calibrated By:

Checked By:

Date:

Date:

Appendix F6

RESULT SHEET

ARTEFACT _____

Measurement point: 70 %rh at 23 °C

Nominal value 70%	Corrected value of Laboratory Standard (%rh)	Corrected value of Laboratory Standard t, (°C)	Indicated value of Artefact (%rh)	Indicated value of Artefact t, (°C)
1				
2				
3				
4				
5				
6				
7				
8				
9				
10				
Average				
Std dev				

Calibrated By:

Checked By:

Date:

Date:

Appendix F7

RESULT SHEET

ARTEFACT _____

Measurement point: 90 %rh at 23 °C

Nominal value 90%	Corrected value of Laboratory Standard (%rh)	Corrected value of Laboratory Standard t, (°C)	Indicated value of Artefact (%rh)	Indicated value of Artefact t, (°C)
1				
2				
3				
4				
5				
6				
7				
8				
9				
10				
Average				
Std dev				

Calibrated By:

Checked By:

Date:

Date:

Template for Uncertainty Budget Table

Quantity	Value	Divisor	Standard Uncertainty	Probability Distribution	Degree of Freedom	Sensitivity Coefficient	Uncertainty Contribution
X_i	x_i		$u(x_i)$		ν_i	c_i	$u_i(y)$

Average Value	= _____ %rh
Relative Humidity Combined Uncertainty, $u_c(y)$	= _____ %rh
Effective degrees of freedom, ν_{eff}	= _____
Coverage factor, k	= _____
Relative Humidity Expanded Uncertainty, $k \times u_c(y)$	= _____ %rh

1. (Add any comments here)

APPENDIX B

CIRCULATION SCHEDULE

Table 1 – Circulation schedule for Artefact A

Economy	Accreditation Body	No. of Lab	Receipt	Dispatch
Malaysia	NML (Ref Lab)	1	-	30 Sept 08
Singapore	Singapore Accreditation Council	2	2 Oct 08	20 Oct 08
Hong Kong	Hong Kong Accreditation Service (HKAS)	2	22 Oct 08	10 Nov 08
Taiwan	Taiwan Accreditation Foundation (TAF)	2	12 Nov 08	2 Dec 08
Malaysia	NML (Ref Lab)	1	6 Dec 08	29 Dec 08
Korea	KOLAS	1	31 Dec 08	no info
China	China National Accreditation Service for Conformity Assessment (CNAS)	2	16 Jan 09	6 Feb 09
Malaysia	NML (Ref Lab)	1	10 Feb 09	20 March 09
Malaysia	STANDARDS MALAYSIA	3	23 March 09	12 April 09
Malaysia	NML (Ref Lab)	1	12 April 09	-

Table 2 – Circulation schedule for Artefact B

Economy	Accreditation Body	No. of Lab	Receipt	Dispatch
Malaysia	NML (Ref Lab)	1	-	30 Sept 08
Australia	National Association of Testing Authorities (NATA)	2	2 Oct 08	24 Oct 08
Indonesia	KOMITE AKREDITASI NASIONAL	2	4 Nov 08	25 Nov 08
Malaysia	NML (Ref Lab)	1	30 Nov 08	5 Dec 08
USA	American Association for Laboratory Accreditation (A2LA)	2	18 Dec 08	no info
USA	National Voluntary Laboratory Accreditation Program	2	no info	no info
USA	International Accreditation Service	2	no info	29 May 09
Malaysia	NML (Ref Lab)	1	22 Jun 09	31 July 09
Malaysia	STANDARDS MALAYSIA	2	31 July 09	28 Aug 09
Malaysia	NML (Ref Lab)	1	30 Aug 09	-

APPENDIX C

REFERENCE VALUE

Table 1 – Reference value 1 for Artefact A

Lab Code: 253, 257, 296, 293, 278, 273

Maximum drift is 1.63 %rh.

Relative Humidity at Nominal (%rh)	Reference Value 1	U_{Ref} without drift	U_{REF} with drift
10	-0.13	0.7	2.0
30	0.21	0.8	2.0
50	1.27	1.0	2.1
70	2.32	1.2	2.2
90	2.64	1.4	2.4

Table 2 – Reference value 2 for Artefact A

Lab Code: 237, 239, 214

Maximum drift is 0.78 %rh

Relative Humidity at Nominal (%rh)	Reference Value 2	U_{Ref} without drift	U_{REF} with drift
10	0.36	0.7	1.1
30	0.92	0.8	1.2
50	1.86	1	1.3
70	2.73	1.2	1.4
90	3.12	1.4	1.6

Table 3 – Reference value 3 for Artefact A

Lab Code: 215, 217, 251

Maximum drift is 1.22 %rh

Relative Humidity at Nominal (%rh)	Reference Value 3	U_{Ref} without drift	U_{REF} with drift
10	0.86	0.7	1.6
30	0.79	0.8	1.6
50	1.73	1	1.7
70	2.87	1.2	1.8
90	3.38	1.4	1.9

Table 4 – Reference value 4 for Artefact B

Lab Code: 266, 260, 267, 261

Maximum drift is 0.86 %rh

Relative Humidity at Nominal (%rh)	Reference Value 4	U_{Ref} without drift	U_{REF} with drift
10	-0.34	0.7	1.2
30	0.40	0.8	1.3
50	1.42	1	1.4
70	2.38	1.2	1.6
90	2.95	1.4	1.8

Table 5 – Reference value 5 for Artefact B

Lab Code: 275, 279, 224

Maximum drift is 1.09 %rh

Relative Humidity at Nominal (%rh)	Reference Value 5	U_{Ref} without drift	U_{REF} with drift
10	-0.1	0.7	1.4
30	1.0	0.8	1.5
50	2.2	1.0	1.6
70	3.2	1.2	1.7
90	3.7	1.4	1.9


Table 6 – Reference value 6 for Artefact B

Lab Code: 271, 263


Maximum drift is 1.3 %rh

Relative Humidity at Nominal (%rh)	Reference Value 6	U_{Ref} without drift	U_{REF} with drift
30	0.9	0.7	1.7
50	2.3	0.9	1.7
70	3.5	1.0	1.8
90	4.1	1.2	1.9

Graph 1: Reference Value for Artefact A



Graph 2: Reference Value for Artefact B



APPENDIX D

PERFORMANCE OF ARTEFACTS BEFORE AND AFTER EACH CALIBRATION

Table 1 – Performance of Artefact A before and after each calibration

Bil	Laboratory Code	Before (%rh)	After (%rh)	Difference (%rh)
1	214	49.9	49.8	0.1
2	215	50.2	50.2	0.0
3	217	51.0	50.2	0.8
4	237	50.2	50.2	0.0
5	239	51.4	51.2	0.2
6	251	51.1	50.4	0.7
7	253	51.2	49.0	2.2
8	257	51.2	49.0	2.2
9	273	50.0	50.0	0.0
10	278	50.0	50.0	0.0
11	293	49.9	50.0	-0.1
12	296	48.6	48.6	0.0
13	Reference Value 1	49.7	49.6	0.1
14	Reference Value 2	49.9	49.9	0.0
15	Reference Value 3	49.6	49.7	-0.1

Table 1 – Performance of Artefact B before and after each calibration

Bil	Laboratory Code	Before (%rh)	After (%rh)	Difference (%rh)
1	224	50.1	50.0	0.1
2	260	49.3	50.4	-1.1
3	261	50.6	50.5	0.1
4	263	49.8	48.3	1.5
5	266	50.0	50.9	-1.0
6	267	49.9	50.1	-0.2
7	271	51.1	49.1	2.0
8	275	50.0	50.0	0.0
9	279	50.0	50.0	0.0
10	Reference Value 4	49.8	50.2	-0.4
11	Reference Value 5	49.3	50.0	-0.7
12	Reference Value 6	49.6	49.6	0.0

APPENDIX E

EXAMPLE OF UNCERTAINTY BUDGET TABLE

CALIBRATION OF RELATIVE HUMIDITY TRANSMITTER

The measurement involves the calibration of Relative Humidity transmitter by comparison method with humidity standard. The measurement was done at nominal values 10 %rh, 30 %rh, 50 %rh, 70 %rh and 90 %rh. at nominal temperature of 23 °C. At each generated condition a time of not less than 60 minutes was be allowed for the relative humidity to equilibrate. A set of 10 readings at 1 minute interval were recorded from the laboratory standard and artefact.

Example for uncertainty components and calculation at 90 %rh is shown in Tables 1.

Table 1 : Uncertainty Budget and Sources of Uncertainty

Symbol	Standard Dew Point Hygrometer	Type	Value	Distribution	Divisor	Std U	Dof
U_a	Calibration	B	0.04	normal	2	0.0200	100000000
U_b	Long term drift	B	0.05	rectangular	1.732	0.0289	100000000
U_c	Reproducibility	B	0.050	normal	1	0.0500	49
U_d	Short term stability	A	0.0045	normal	1	0.0045	9
U_e	Non-linearity	B	0.016	rectangular	1.732	0.0092	100000000
U_f	Settling time to final value	B	0.01	rectangular	1.732	0.0058	100000000
U_g	Resolution	B	0.005	rectangular	1.732	0.0029	100000000
U_h	Effect of pressure difference	B	0.05	rectangular	1.732	0.0289	100000000
U_i	Error due to condition different from calibration	B	0.05	rectangular	1.732	0.0289	100000000
U_j	Hysteresis	B	0.05	rectangular	1.732	0.0289	100000000
U_k	Contamination	B	0.05	rectangular	1.732	0.0289	100000000
	Combined uncertainty in dew point					0.0849	408
U_{dp}	Combined uncertainty in Relative Humidity					-0.4700	
	Temperature Standard						
	NML-356/05						
U_a	Calibration	B	0.01	normal	2	0.0050	100000000
U_b	Long term drift	B	0	rectangular	1.732	0.0000	100000000
U_c	Reproducibility	B	0.05	normal	1	0.0500	49
U_d	Short term stability	A	0.0016	normal	1	0.0016	9
U_e	Non-linearity	B	0.008	rectangular	1.732	0.0046	100000000
U_f	Self Heating effect	B	0.025	rectangular	1.732	0.0144	100000000
U_g	Hysteresis	B	0.0000	rectangular	1.732	0.0000	100000000
U_h	Stem Conduction	B	0.01	rectangular	1.732	0.0058	100000000
U_i	Resolution	B	0.005	rectangular	1.732	0.0029	100000000
	Combined uncertainty in temperature					0.0529	61
U_t	Combined uncertainty in Relative Humidity					0.2900	

	Chamber						
U_a	RH uniformity		0.2	rectangular	1.732	0.1155	100000000
U_b	RH stability		0.0218	normal	1	0.0218	9
U_c	Incomplete settling to final rh value		0.2	rectangular	1.732	0.1155	100000000
$U_{chamber}$	Combined uncertainty					0.1648	29361

	Artefact A						
U_a	Standard deviation		0.0000	normal	1	0.0000	9
U_b	Resolution		0.05	rectangular	1.732	0.0289	100000000
U_c	Incomplete settling to final rh value		0.5	rectangular	1.732	0.2887	100000000
U_d	Transpotation drift		1.22	rectangular	1.732	0.7044	100000000
U_{UUT}	Combined uncertainty					0.7618	133052880

U_c	Combined Uncertainty of calibration					0.9552	6960
k	Expanded uncertainty %rh			normal	2	1.9105	infiniti

----- END -----